

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 12/11/2019

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:44
INTERNAL STANDARD	VERIFIED	15:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:44
BLANK TEST	0.00	15:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:45
BLANK TEST	0.00	15:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:46
BLANK TEST	0.00	15:47

