

SUPERVISOR TEST

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 12/25/2019

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:56
INTERNAL STANDARD	VERIFIED	17:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.08	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:59
BLANK TEST	0.00	18:00

