

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 12/25/2019

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:10
INTERNAL STANDARD	VERIFIED	15:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.08	15:10
BLANK TEST	0.00	15:11
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.08	15:11
BLANK TEST	0.00	15:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.08	15:12
BLANK TEST	0.00	15:13

