

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 03/11/2020

OPERATOR NAME:
CHRISTOPHER J DELONG
Permit #: 47017
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 23:28 |
| INTERNAL STANDARD | VERIFIED | 23:28 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 23:28 |
| BLANK TEST | 0.00 | 23:29 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 23:30 |
| BLANK TEST | 0.00 | 23:31 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 23:31 |
| BLANK TEST | 0.00 | 23:32 |

