

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 04/15/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:58
INTERNAL STANDARD	VERIFIED	14:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:58
BLANK TEST	0.00	14:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:59
BLANK TEST	0.00	15:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:01
BLANK TEST	0.00	15:02

