

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906




Date: 03/01/2020

OPERATOR NAME:
EDWARD F CONSTANTINE
Permit #: 45980
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:15
INTERNAL STANDARD	VERIFIED	18:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:16
BLANK TEST	0.00	18:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:17
BLANK TEST	0.00	18:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:18
BLANK TEST	0.00	18:19

BFO



SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 03/04/2020

OPERATOR NAME:
EDWARD F CONSTANTINE
Permit #: 45980
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:04
INTERNAL STANDARD	VERIFIED	20:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:04
BLANK TEST	0.00	20:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:05
BLANK TEST	0.00	20:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:07



BFO

SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 03/11/2020

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:20
INTERNAL STANDARD	VERIFIED	22:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:21
BLANK TEST	0.00	22:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:22
BLANK TEST	0.00	22:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:23
BLANK TEST	0.00	22:24



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 03/18/2020

OPERATOR NAME:
EDWARD F CONSTANTINE
Permit #: 45980
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:12
INTERNAL STANDARD	VERIFIED	18:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:12
BLANK TEST	0.00	18:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:13
BLANK TEST	0.00	18:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:15
BLANK TEST	0.00	18:15

BFC

