

SOLUTION CHANGE

LEB



ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 06/02/2020

OPERATOR NAME:  
BRETT A GRABELDINGER  
Permit #: 46896  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:56
INTERNAL STANDARD	VERIFIED	05:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:56
BLANK TEST	0.00	05:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:58
BLANK TEST	0.00	05:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:59
BLANK TEST	0.00	06:00



SUPERVISOR TEST

LEB  
LATIC

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 06/03/2020

OPERATOR NAME:  
EMRAH LATIC  
Permit #: 46908  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:06
INTERNAL STANDARD	VERIFIED	23:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:07
BLANK TEST	0.00	23:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:08
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:09
BLANK TEST	0.00	23:10

