

SOLUTION CHANGE



ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 10/09/2019

OPERATOR NAME:
EDWARD F CONSTANTINE
Permit #: 45980
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:49
INTERNAL STANDARD	VERIFIED	00:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:49
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:53



SUPERVISOR TEST


ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 10/16/2019

OPERATOR NAME:
EDWARD F CONSTANTINE
Permit #: 45980
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:55
INTERNAL STANDARD	VERIFIED	19:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:55
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:57
BLANK TEST	0.00	19:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:58
BLANK TEST	0.00	19:59



SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 10/31/2019

OPERATOR NAME:
LAUREN C MARLEAU
Permit #: 45975
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:58
INTERNAL STANDARD	VERIFIED	05:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	05:59
BLANK TEST	0.00	05:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:00
BLANK TEST	0.00	06:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:01
BLANK TEST	0.00	06:02



SUPERVISOR TEST

LANC

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

BFO


Date: 11/07/2019

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:41
INTERNAL STANDARD	VERIFIED	02:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:42

Blank Error



SUPERVISOR TEST

Latic
BFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/07/2019

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:45
INTERNAL STANDARD	VERIFIED	02:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:45
BLANK TEST	0.00	02:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:48
BLANK TEST	0.00	02:49



SUPERVISOR TEST

BFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/13/2019

OPERATOR NAME:
LAUREN C MARLEAU
Permit #: 45976
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:41
INTERNAL STANDARD	VERIFIED	15:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:41
BLANK TEST	0.00	15:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:42
BLANK TEST	0.00	15:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:44
BLANK TEST	0.00	15:45



SUPERVISOR TEST

LATIC
BFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/20/2019

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:28
INTERNAL STANDARD	VERIFIED	19:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:28
BLANK TEST	0.00	19:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:31
BLANK TEST	0.00	19:32



SUPERVISOR TEST

BFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 11/27/2019

OPERATOR NAME:
EDWARD F CONSTANTINE
Permit #: 45980
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:19
INTERNAL STANDARD	VERIFIED	21:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:19
BLANK TEST	0.00	21:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:21
BLANK TEST	0.00	21:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:22
BLANK TEST	0.00	21:23



SOLUTION CHANGE

BFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/03/2019

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:42
INTERNAL STANDARD	VERIFIED	01:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:42
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:43
BLANK TEST	0.00	01:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:45
BLANK TEST	0.00	01:46



SUPERVISOR TEST

BFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/05/2019

OPERATOR NAME:
MICHAEL P BAKER
Permit #: 44230
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:41
INTERNAL STANDARD	VERIFIED	00:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:43
BLANK TEST	0.00	00:44



SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/11/2019

OPERATOR NAME:
LAUREN C MARLEAU
Permit #: 45976
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:31
INTERNAL STANDARD	VERIFIED	14:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:31
BLANK TEST	0.00	14:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:32
BLANK TEST	0.00	14:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:33
BLANK TEST	0.00	14:34



SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/21/2019

OPERATOR NAME:
LAUREN C MARLEAU
Permit #: 45976
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:35
INTERNAL STANDARD	VERIFIED	12:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:35
BLANK TEST	0.00	12:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:36
BLANK TEST	0.00	12:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:38
BLANK TEST	0.00	12:38



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906



BFO

Date: 12/25/2019

OPERATOR NAME:
LAUREN C MARLEAU
Permit #: 45976
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:33
INTERNAL STANDARD	VERIFIED	14:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:33
BLANK TEST	0.00	14:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:35
BLANK TEST	0.00	14:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:36
BLANK TEST	0.00	14:37



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 12/25/2019

OPERATOR NAME:
LAUREN C MARLEAU
Permit #: 45976
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:33
INTERNAL STANDARD	VERIFIED	14:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:33
BLANK TEST	0.00	14:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:35
BLANK TEST	0.00	14:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:36
BLANK TEST	0.00	14:37

