

BFO

SUPERVISOR TEST



ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 04/22/2020

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:29
INTERNAL STANDARD	VERIFIED	23:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:29
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:31
BLANK TEST	0.00	23:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:32
BLANK TEST	0.00	23:33



SUPERVISOR TEST

BFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 04/29/2020

OPERATOR NAME:  
BRETT A GRABELDINGER  
Permit #: 46896  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:39
INTERNAL STANDARD	VERIFIED	22:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:40
BLANK TEST	0.00	22:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:42
BLANK TEST	0.00	22:43

