

SUPERVISOR TEST

-----  
 SHERRILL CITY  
 POLICE DEPT  
 DATAMASTER dmt: 115108  
 -----

Date: 04/22/2020

OPERATOR NAME:  
 ROBERT J DRAKE  
 Permit #: 27298  
 Operator Agency/Dept: SHERRILL CITY PD  
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:53
INTERNAL STANDARD	VERIFIED	06:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:54
BLANK TEST	0.00	06:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:56
BLANK TEST	0.00	06:57

