

SUPERVISOR TEST

-----  
 SHERRILL CITY  
 POLICE DEPT  
 DATAMASTER dmt: 115108  
 -----

Date: 06/10/2020

OPERATOR NAME:  
 ROBERT J DRAKE  
 Permit #: 27298  
 Operator Agency/Dept: SHERRILL CITY PD  
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:18
INTERNAL STANDARD	VERIFIED	07:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:20
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22

