

SUPERVISOR TEST

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 06/24/2020

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:58
INTERNAL STANDARD	VERIFIED	06:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:59
BLANK TEST	0.00	07:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:00
BLANK TEST	0.00	07:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:02

