

```

<?xml version="1.0" standalone="yes"?>
<TRACS_IMPORT>
  <ALCOTEST_VALUES>
    <SERIAL>ARAH-0026</SERIAL>
    <IP_ADRESS>10.81.201.28</IP_ADRESS>
    <DATE>07/12/2020</DATE>
    <TIME>05:57PM</TIME>
  </ALCOTEST_VALUES>
  <CALCHECK_1>
    <CAL_MESSUNG_2_MESSUNG_ERROR>0</CAL_MESSUNG_2_MESSUNG_ERROR>
    <CAL_MESSUNG_2_IR_AAK_NULLWERT_1>0.00</CAL_MESSUNG_2
_IR_AAK_NULLWERT_1>
    <CAL_MESSUNG_2_IR_AAK_MESSWERT>0.10</CAL_MESSUNG_2
_IR_AAK_MESSWERT>
    <CAL_MESSUNG_2_EC_AAK_MESSWERT>0.09</CAL_MESSUNG_2
_EC_AAK_MESSWERT>
    <CAL_MESSUNG_2_IR_AAK_NULLWERT_2>0.00</CAL_MESSUNG_2
_IR_AAK_NULLWERT_2>
    <ID_BLANK_TIME_BEFORE_2>05:42:13PM
07/12/2020</ID_BLANK_TIME_BEFORE_2>
    <ID_TIME_START_BLOW_2>--:--:-- --/--/-----
</ID_TIME_START_BLOW_2>
    <ID_TIME_MEAS_2>05:42:50PM 07/12/2020</ID_TIME_MEAS_2>
    <ID_BLANK_TIME_AFTER_2>05:43:23PM
07/12/2020</ID_BLANK_TIME_AFTER_2>
  </CALCHECK_1>
  <SUBJECT_TEST>
    <PROB_MESSUNG_2_MESSUNG_ERROR>0</PROB_MESSUNG_2
_MESSUNG_ERROR>
    <PROB_MESSUNG_2_IR_AAK_NULLWERT_1>0.00</PROB_MESSUNG_2
_IR_AAK_NULLWERT_1>
    <PROB_MESSUNG_2_IR_AAK_MESSWERT>0.27</PROB_MESSUNG_2
_IR_AAK_MESSWERT>
    <PROB_MESSUNG_2_EC_AAK_MESSWERT>0.28</PROB_MESSUNG_2
_EC_AAK_MESSWERT>
    <PROB_MESSUNG_2_IR_AAK_NULLWERT_2>0.00</PROB_MESSUNG_2
_IR_AAK_NULLWERT_2>
    <PROB_MESSUNG_2_BLAS_FEHLVERSUCHE>2</PROB_MESSUNG_2
_BLAS_FEHLVERSUCHE>
    <PROB_MESSUNG_2_BLAS_VOLUMEN>2.7</PROB_MESSUNG_2
_BLAS_VOLUMEN>
    <PROB_MESSUNG_2_BLAS_ZEIT>23.3</PROB_MESSUNG_2_BLAS_ZEIT>
    <ID_BLANK_TIME_BEFORE_3>05:49:28PM 07/12/2020,05:44:31PM
07/12/2020,05:47:10PM 07/12/2020</ID_BLANK_TIME_BEFORE_3>
    <ID_TIME_START_BLOW_3>05:50:03PM 07/12/2020,05:45:32PM
07/12/2020,05:47:50PM 07/12/2020</ID_TIME_START_BLOW_3>
    <ID_TIME_END_BLOW_3>05:50:27PM 07/12/2020,05:45:38PM
07/12/2020,05:48:01PM 07/12/2020</ID_TIME_END_BLOW_3>
    <ID_TIME_MEAS_3>05:50:35PM 07/12/2020,05:45:38PM
07/12/2020,05:48:11PM 07/12/2020</ID_TIME_MEAS_3>
    <ID_BLANK_TIME_AFTER_3>05:52:02PM
07/12/2020</ID_BLANK_TIME_AFTER_3>

```

```

</SUBJECT_TEST>
<CALCHECK_2>
  <CAL_MESSUNG_3_MESSUNG_ERROR>0</CAL_MESSUNG_3_MESSUNG_ERROR>
  <CAL_MESSUNG_3_IR_AAK_NULLWERT_1>0.00</CAL_MESSUNG_3
_IR_AAK_NULLWERT_1>
  <CAL_MESSUNG_3_IR_AAK_MESSWERT>0.09</CAL_MESSUNG_3
_IR_AAK_MESSWERT>
  <CAL_MESSUNG_3_EC_AAK_MESSWERT>0.09</CAL_MESSUNG_3
_EC_AAK_MESSWERT>
  <CAL_MESSUNG_3_IR_AAK_NULLWERT_2>0.00</CAL_MESSUNG_3
_IR_AAK_NULLWERT_2>
  <ID_BLANK_TIME_BEFORE_4>05:52:02PM
07/12/2020</ID_BLANK_TIME_BEFORE_4>
  <ID_TIME_START_BLOW_4>--:--:-- --/--/-----
</ID_TIME_START_BLOW_4>
  <ID_TIME_MEAS_4>05:52:04PM 07/12/2020</ID_TIME_MEAS_4>
  <ID_BLANK_TIME_AFTER_4>05:52:37PM
07/12/2020</ID_BLANK_TIME_AFTER_4>
</CALCHECK_2>
<GENERAL_DATA>
  <ID_SERIAL_NO>ARAH-0026</ID_SERIAL_NO>
  <ID_SW_VERSION> 8321288 1.12.2</ID_SW_VERSION>
  <ID_START_TIME>05:37:35PM 07/12/2020</ID_START_TIME>
  <ID_END_TIME>05:53:31PM 07/12/2020</ID_END_TIME>
  <ID_LAST_CAL_DATE>05/22/2020</ID_LAST_CAL_DATE>
  <ID_AAK_ENDRESULT> 0.27 %BAC</ID_AAK_ENDRESULT>
  <ID_TEST_COUNTER>536</ID_TEST_COUNTER>
  <ID_MEAS_PROC>1</ID_MEAS_PROC>
  <ID_CALIBRATION_COUNTER>104</ID_CALIBRATION_COUNTER>
  <ID_ERROR_PROTOKOLL>0</ID_ERROR_PROTOKOLL>
  <ID_TEST_GAS_TYPE>2</ID_TEST_GAS_TYPE>
  <ID_TEST_GAS_INLET>1</ID_TEST_GAS_INLET>
  <ID_TEST_GAS_CONC>2.00</ID_TEST_GAS_CONC>
  <ID_TEST_CFG_SENSOR>0</ID_TEST_CFG_SENSOR>
  <ID_DATA_INPUT_ON_OFF>1</ID_DATA_INPUT_ON_OFF>
  <ID_IR_AVERAGE>20653</ID_IR_AVERAGE>
  <ID_EC_SENSOR_OFFSET>113</ID_EC_SENSOR_OFFSET>
  <ID_HOSE_TEMP>47.0</ID_HOSE_TEMP>
  <ID_IR_SENSOR_TEMP>45.3</ID_IR_SENSOR_TEMP>
  <ID_INTERNAL_DEVICE_TEMP>38.1</ID_INTERNAL_DEVICE_TEMP>
  <ID_12V_VOLTAGE>12.3</ID_12V_VOLTAGE>
  <ID_3V_BACKUP_VOLTAGE>3.3</ID_3V_BACKUP_VOLTAGE>
  <ID_MS_SW_CRC>0x6581</ID_MS_SW_CRC>
  <ID_PRESSURE_DRY_GAS>0</ID_PRESSURE_DRY_GAS>
  <ID_DRY_GAS_LOT_NO>1057590 CYL73</ID_DRY_GAS_LOT_NO>
  <ID_DRY_GAS_EXP_DATA>12/08/2021</ID_DRY_GAS_EXP_DATA>
</GENERAL_DATA>
<INPUT_VALUES>
  <UF_CASE_NUMBER>9700807</UF_CASE_NUMBER>
  <ZERO_TOLERANCE_CASE>NO</ZERO_TOLERANCE_CASE>
  <TROOP>T</TROOP>
  <STATION>Junius Ponds</STATION>

```

<ARREST\_DATE>07/12/2020</ARREST\_DATE>  
<ARREST\_TIME>05:03PM</ARREST\_TIME>  
<ARRESTING\_MEM\_LAST>rondon</ARRESTING\_MEM\_LAST>  
<ARRESTING\_MEM\_FIRST>jeanette</ARRESTING\_MEM\_FIRST>  
<ARRESTING\_MEM\_MI>m</ARRESTING\_MEM\_MI>  
<ARRESTING\_MEM\_SHIELD>1343</ARRESTING\_MEM\_SHIELD>  
<SUBJECT\_LAST\_NAME>suko</SUBJECT\_LAST\_NAME>  
<SUBJECT\_FIRST\_NAME>sedina</SUBJECT\_FIRST\_NAME>  
<SUBJECT\_MI>-</SUBJECT\_MI>  
<SUBJECT\_DOB>01/29/1987</SUBJECT\_DOB>  
<SUBJECT\_AGE>33</SUBJECT\_AGE>  
<SUBJECT\_GENDER>FEMALE</SUBJECT\_GENDER>  
<OPERATOR\_LAST\_NAME>rondon</OPERATOR\_LAST\_NAME>  
<OPERATOR\_FIRST\_NAME>jeanette</OPERATOR\_FIRST\_NAME>  
<OPERATOR\_MI>m</OPERATOR\_MI>  
<OPERATOR\_SHIELD\_NO>1343</OPERATOR\_SHIELD\_NO>  
<OPERATOR\_PERMIT\_NO>46439</OPERATOR\_PERMIT\_NO>  
<OPERATOR\_PERMIT\_EXP>11/24/2021</OPERATOR\_PERMIT\_EXP>  
<SUBJECT\_OBSERVED\_20\_MINUTES>YES</SUBJECT\_OBSERVED\_20  
\_MINUTES>  
<TEST\_20\_MINUTES\_PAST\_ARREST>YES</TEST\_20  
\_MINUTES\_PAST\_ARREST>  
<REFUSAL\_MIRANDA\_READ>YES</REFUSAL\_MIRANDA\_READ>  
</INPUT\_VALUES>  
</TRACS\_IMPORT>