

**Preliminary Breath Test Device
Monthly Verification Log**

Department NYMPD

Device Serial # 500475

Test Date	Sim. Solution Lot #	Sim. Temperature			Sim. Serial #	Test Results			Officer Performing Verification	Sim. Solution Change *
		1	2	3		1	2	3		

10/15/19	19180	33.99	33.99	33.99	MP5218	.09	.09	.09	Frankland	✓
10/16/19	19180	33.98	33.99	33.99	MP5218	.09	.09	.09	Frankland	
10/22/19	19180	33.99	33.99	33.98	MP5218	.09	.09	.09	Frankland	
10/23/19	19180	33.99	33.99	33.99	MP5218	.09	.09	.09	Frankland	
10/30/19	19180	33.99	33.99	33.99	MP5218	.09	.09	.09	Frankland	
10/31/19	19180	33.99	33.99	33.99	MP5218	.09	.09	.09	Frankland	
11/6/19	19180	33.99	33.99	33.99	MP5218	.09	.09	.09	Frankland	
11/13/19	19180	33.99	33.99	33.99	MP5218	.08	.08	.08	Frankland ^{Out of Service}	DCS contacted
11/15/19	19180	33.99	33.99	33.99	MP5218	.10	.10	.10	Frankland ^{In Service}	New Solution
11/20/19	19180	33.99	33.99	33.99	MP5218	.10	.10	.10	Frankland	
11/27/19	19180	33.99	33.99	33.99	MP5218	.10	.10	.10	Frankland	
11/30/19	19250	33.99	33.99	33.99	MP5218	.09	.09	.09	Frankland	✓
12/4/19	19250	33.99	33.99	33.99	MP5218	.09	.09	.09	Frankland	
12/11/19	19250	33.99	33.99	33.99	MP5218	.09	.09	.09	Frankland	

* - Please enter a ✓ to denote a simulator solution lot number change

Version 8/12/2015