

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 02/29/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:09
INTERNAL STANDARD	VERIFIED	15:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:09
BLANK TEST	0.00	15:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:10
BLANK TEST	0.00	15:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:11
BLANK TEST	0.00	15:12

