

SUPERVISOR TEST

NEW YORK MILLS
POLICE DEPT.
DATAMASTER dmt: 500475

Date: 01/01/2020

OPERATOR NAME:
ROBERT C FRANKLAND
Permit #: 40674
Operator Agency/Dept: NY MILLS PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:50
INTERNAL STANDARD	VERIFIED	16:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:50
BLANK TEST	0.00	16:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:52
BLANK TEST	0.00	16:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:53
BLANK TEST	0.00	16:54

