

**SUPERVISOR TEST**

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 11/15/2019

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:38
INTERNAL STANDARD	VERIFIED	16:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:38
BLANK TEST	0.00	16:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:39
BLANK TEST	0.00	16:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:41
BLANK TEST	0.00	16:41

