

SUPERVISOR TEST

NEW YORK MILLS  
POLICE DEPT.  
DATAMASTER dmt: 500475

Date: 10/22/2019

OPERATOR NAME:  
ROBERT C FRANKLAND  
Permit #: 40674  
Operator Agency/Dept: NY MILLS PD  
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:21
INTERNAL STANDARD	VERIFIED	19:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:22
BLANK TEST	0.00	19:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25

