

**Infrared Instrument  
Weekly Calibration Verification Test Record**

Department: \_\_\_\_\_

Instrument Serial #: \_\_\_\_\_

OC 50 - WFO

100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
11-7-19	19180	34.02 34.02	34.02	34.02	MP4192	.10	.10	.10	LATIC	NO
11/13/19	19180 53215	34.03	34.02	34.03	MP4192	.10	.10	.10	Dillenbeck	NO
11/26/19	19180	34.02	34.03	34.03	MP4192	.10	.10	.10	LATIC	NO
11/27/19	19180	34.02	34.02	34.02	MP4192	.10	.10	.10	Gibbons	No
12/3/19	19250	34.02	34.03	34.03	MP4192	.10	.10	.10	Grabbe	YES
12/6/19	19250	34.02	34.02	34.02	MP4192	.10	.10	.10	LATIC	NO
12/11/19	19250	34.03	34.03	34.02	MP4192	.10	.10	.10	Gibbons	No
12/18/19	19250	34.02	34.02	34.02	MP4192	.10	.10	.10	LATIC	NO
12/26/19	19250	34.03	34.03	34.03	MP4192	.10	.10	.10	Gibbons	NO
01/05/20	19250	34.02	34.03	34.03	MP4192	.10	.10	.10	Gibbons	NO
01/08/20	19250	34.02	34.03	34.03	MP4192	.10	.10	.10	Gibbons	No
1/16/20	19250	34.03	34.03	34.02	MP4192	.10	.10	.10	LATIC	NO

\* - Please enter a √ to denote simulator solution change taking place