

SUPERVISOR TEST

WFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/16/2019

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:24
INTERNAL STANDARD	VERIFIED	23:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:25
BLANK TEST	0.00	23:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:26
BLANK TEST	0.00	23:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:28
BLANK TEST	0.00	23:28



SUPERVISOR TEST

LATIC

WFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/07/2019

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:17
INTERNAL STANDARD	VERIFIED	00:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:19
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:20
BLANK TEST	0.00	00:21



SUPERVISOR TEST

WFO




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/13/2019

OPERATOR NAME:
LEONARD F DILLENBECK
Permit #: 47621
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:11
INTERNAL STANDARD	VERIFIED	21:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:12
BLANK TEST	0.00	21:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:13
BLANK TEST	0.00	21:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	21:14
BLANK TEST	0.00	21:15



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




Date: 10/23/2019

OPERATOR NAME:
MICHAEL P GIBBONS
Permit #: 45972
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:19
INTERNAL STANDARD	VERIFIED	01:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:22
BLANK TEST	0.00	01:23

WFO
Weekly Test



SUPERVISOR TEST

LATIC

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




WFO

Date: 11/20/2019

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:05
INTERNAL STANDARD	VERIFIED	20:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:06
BLANK TEST	0.00	20:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:08
BLANK TEST	0.00	20:09



SOLUTION CHANGE

WFO




ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

Date: 12/03/2019

OPERATOR NAME:
 BRETT A GRABELDINGER
 Permit #: 46896
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:38
INTERNAL STANDARD	VERIFIED	22:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:40
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:42



SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




LATIC

Date: 12/05/2019

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:06
INTERNAL STANDARD	VERIFIED	01:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:07
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:09
BLANK TEST	0.00	01:10



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/27/2019




OPERATOR NAME:
MICHAEL P GIBBONS
Permit #: 45972
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:22
INTERNAL STANDARD	VERIFIED	22:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:23
BLANK TEST	0.00	22:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:24
BLANK TEST	0.00	22:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:25
BLANK TEST	0.00	22:26

WFO weekly

Datamaster



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197




Date: 12/11/2019

OPERATOR NAME:
MICHAEL P GIBBONS
Permit #: 45972
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:35
INTERNAL STANDARD	VERIFIED	20:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:35
BLANK TEST	0.00	20:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:36
BLANK TEST	0.00	20:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:37
BLANK TEST	0.00	20:38

WFO Weekly
Datamaster



SUPERVISOR TEST

NFO
LATIC

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/18/2019

OPERATOR NAME:
EMRAH LATIC
Permit #: 46908
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:29
INTERNAL STANDARD	VERIFIED	21:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:29
BLANK TEST	0.00	21:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:31
BLANK TEST	0.00	21:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:32
BLANK TEST	0.00	21:33

