

SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606




*Handwritten mark*

Date: 05/20/2020

OPERATOR NAME:  
EMRAH LATIC  
Permit #: 46908  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:16
INTERNAL STANDARD	VERIFIED	23:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:16
BLANK TEST	0.00	23:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:17
BLANK TEST	0.00	23:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:19
BLANK TEST	0.00	23:19



SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 05/27/2020

OPERATOR NAME:  
BRETT A GRABELDINGER  
Permit #: 46896  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:58
INTERNAL STANDARD	VERIFIED	17:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:59
BLANK TEST	0.00	18:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:00
BLANK TEST	0.00	18:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:01
BLANK TEST	0.00	18:02

