

SUPERVISOR TEST

CFO




ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 05/27/2020

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:08
INTERNAL STANDARD	VERIFIED	22:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:10
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:11
BLANK TEST	0.00	22:12



SOLUTION CHANGE

CF0




ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/02/2020

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:08
INTERNAL STANDARD	VERIFIED	02:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:09
BLANK TEST	0.00	02:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:11
BLANK TEST	0.00	02:12



SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
DATAMASTER dmt: 123906

Date: 06/04/2020

OPERATOR NAME:
BRANDON J RUTLEDGE
Permit #: 43762
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:18
INTERNAL STANDARD	VERIFIED	03:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:18
BLANK TEST	0.00	03:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:19
BLANK TEST	0.00	03:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:21
BLANK TEST	0.00	03:22

CFO

