

SUPERVISOR TEST

BFO



ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/05/2020

OPERATOR NAME:
BRETT A GRABELDINGER
Permit #: 46896
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:48
INTERNAL STANDARD	VERIFIED	23:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:50
BLANK TEST	0.00	23:51



SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 141906

Date: 01/08/2020

OPERATOR NAME:
LAUREN C MARLEAU
Permit #: 45976
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:28
INTERNAL STANDARD	VERIFIED	19:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:28
BLANK TEST	0.00	19:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:31
BLANK TEST	0.00	19:32

