

BFO

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 141906

Date: 01/15/2020

OPERATOR NAME:  
LEONARD F DILLENBECK  
Permit #: 47621  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:46
INTERNAL STANDARD	VERIFIED	21:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:47
BLANK TEST	0.00	21:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:48
BLANK TEST	0.00	21:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:49
BLANK TEST	0.00	21:50

