

SOLUTION CHANGE




ROME CITY
POLICE DEPT
DATAMASTER dmt: 106106

Date: 10/01/2019

OPERATOR NAME:
HOLLIE B SILVERMAN
Permit #: 39044
Operator Agency/Dept: ROME CITY PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:48
INTERNAL STANDARD	VERIFIED	19:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:49
BLANK TEST	0.00	19:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:50
BLANK TEST	0.00	19:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:51
BLANK TEST	0.00	19:52



Hollie B. Silverman Rome PD 10/01/19
Hollie Silverman 10805

SUPERVISOR TEST




ROME CITY
POLICE DEPT
DATAMASTER dmt: 106106

Date: 10/04/2019

OPERATOR NAME:
HOLLIE B SILVERMAN
Permit #: 39044
Operator Agency/Dept: ROME CITY PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:28
INTERNAL STANDARD	VERIFIED	15:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:29
BLANK TEST	0.00	15:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:30
BLANK TEST	0.00	15:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:31
BLANK TEST	0.00	15:32



Hollie Silverman Rome PD 10/04/19
Hollie Silverman WKS

SUPERVISOR TEST




ROME CITY
POLICE DEPT
DATAMASTER dmt: 106106

Date: 10/11/2019

OPERATOR NAME:
HOLLIE B SILVERMAN
Permit #: 39044
Operator Agency/Dept: ROME CITY PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:24
INTERNAL STANDARD	VERIFIED	18:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:24
BLANK TEST	0.00	18:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:26
BLANK TEST	0.00	18:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:27
BLANK TEST	0.00	18:28



Hollie Silverman Rome PD 10/11/19
Hollie Silverman *10/11/19*

SUPERVISOR TEST




ROME CITY
POLICE DEPT
DATAMASTER dmt: 106106

Date: 10/18/2019

OPERATOR NAME:
HOLLIE B SILVERMAN
Permit #: 39044
Operator Agency/Dept: ROME CITY PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:41
INTERNAL STANDARD	VERIFIED	11:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:42
BLANK TEST	0.00	11:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:43
BLANK TEST	0.00	11:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:44
BLANK TEST	0.00	11:45



Hollie Silverman Rome PD 10/18/19
Hollie Silverman 10905

SUPERVISOR TEST




ROME CITY
POLICE DEPT
DATAMASTER dmt: 106106

Date: 10/25/2019

OPERATOR NAME:
HOLLIE B SILVERMAN
Permit #: 39044
Operator Agency/Dept: ROME CITY PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:36
INTERNAL STANDARD	VERIFIED	17:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:36
BLANK TEST	0.00	17:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:37
BLANK TEST	0.00	17:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:39
BLANK TEST	0.00	17:39



Hollie B Silverman ROME PD 10/25/19
Hollie Silverman 10905