

SUPERVISOR TEST

ROME CITY
POLICE DEPT
DATAMASTER dmt: 106106

Date: 01/24/2020

OPERATOR NAME:
HOLLIE B SILVERMAN
Permit #: 39044
Operator Agency/Dept: ROME CITY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:41
INTERNAL STANDARD	VERIFIED	19:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:41
BLANK TEST	0.00	19:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:43
BLANK TEST	0.00	19:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:44
BLANK TEST	0.00	19:45

