

SUPERVISOR TEST

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 04/08/2020

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:01
INTERNAL STANDARD	VERIFIED	07:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04

