

SUPERVISOR TEST

SHERRILL CITY
POLICE DEPT
DATAMASTER dmt: 115108

Date: 05/28/2020

OPERATOR NAME:
ROBERT J DRAKE
Permit #: 27298
Operator Agency/Dept: SHERRILL CITY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:53
INTERNAL STANDARD	VERIFIED	06:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:53
BLANK TEST	0.00	06:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:56
BLANK TEST	0.00	06:57

