

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 06/12/2020

RMS
20-20588

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:34
INTERNAL STANDARD	VERIFIED	15:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:34
BLANK TEST	0.00	15:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:35
BLANK TEST	0.00	15:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:37
BLANK TEST	0.00	15:37

